



**LUMEN DYNAMICS**

PUTTING YOU IN CONTROL

An Excelitas Technologies® Company



## CE Declaration of Conformity

**Manufacturer's Name:** Lumen Dynamics Group Inc.  
**Manufacturer's Address:** 2260 Argentia Road, Mississauga, Ontario L5N 6H7 Canada

**Authorized EU Representative:** Excelitas Technologies Elcos GmbH  
Hans-Riedl-Str. 9, D-85622 Feldkirchen, Germany  
Tel: (+49) 89-255 458-0/ (+49) 89-255 458-111

**Declares that the product:** **Equipment Type:** Laser-LED Microscope Illumination Device  
**Trade Name/Model No.:** X-Cite Turbo/ XT600-T

**Conforms to the following Council Directives:**

<b>Low Voltage</b>	<b>2014/35/EU</b>
<b>EMC</b>	<b>2014/30/EU</b>
<b>WEEE</b>	<b>2012/19/EU</b>
<b>RoHS</b>	<b>2011/65/EU</b>

### **Conforms to the following Standards:**

#### **Product Safety**

CAN/CSA C22.2 No. 61010-1:2012	Safety Requirements for Electrical Equipment for Measurement, Control and Laboratory Use, Part 1: General Requirements
UL 61010-1:2012 / EN61010-1:2010	
IEC 61010-1:2010 (Third Edition)	Safety Requirements for Electrical Equipment for Measurement, Control and Laboratory Use, Part 1: General Requirements
IEC 62471:2006 (First Edition), including:	Photobiological safety of lamps and lamp systems
IEC 60825-1:2014	Safety of Laser Products – Part 1: Equipment Classification and Requirements 2014-05
21 CFR 1040.10 and 1040.11	FDA/CDRH Code of Regulations, Title 21 – Food and Drugs, Section 1040.10 – Laser Products and 1040.11 – Specific Purpose Laser Products, except for deviations pursuant to Laser Notice No. 50, dated June 24, 2007

#### **Electromagnetic Compatibility**

IEC 61326-1:2012 / EN 61326-1:2013	Electrical Equipment for Measurement, Control and Laboratory Use – EMC Requirements – Part 1 General Requirements
FCC Part 15 Subpart-B: 2015	Code of Federal Regulations – Radio Frequency Devices
ICES-003: Issue 5:2012	Information Technology Equipment (ITE) – Limits and Methods of Measurement

I, the undersigned, hereby declare that the equipment specified above conforms to the above directives and standards specified.

Signature 

SEP 08 2016  
Date

**Full Name:** Jim Farrell  
**Position:** QA, QMA & EMS Manager

**X-Cite®**  
Fluorescence Illumination • In Control

**OmniCure®**  
UV Curing • In Control

www.excelitas.com/x-cite  
x-cite@excelitas.com

www.excelitas.com/omnicure  
omnicure@excelitas.com

Lumen Dynamics Group Inc. 2260 Argentia Road, Mississauga, Ontario, L5N 6H7 CANADA  
Telephone: +1 905 821-2600 • Toll Free (USA and CAN): +1 800 668-8752 • Fax: +1 905 821-2055